



Plumsted Township School District

117 Evergreen Road
New Egypt, NJ 08533
Telephone (609) 758-6800 Ext. 4206
Fax (609) 758-6808
www.newegypt.us

Gerald North
Superintendent of Schools

Sean Gately
Business Administrator/ Board Secretary

UPDATE

April 3, 2017

Dear Parent/Guardian:

This letter is to inform you that there has been a recall of epi-pens 0.3 mg due to failure to activate or requiring excessive force to activate posing a very serious, life-threatening issue to those suffering from serious allergies.

If your child is identified as a student with a severe allergy requiring the use of an epi-pen, the following batch numbers were confirmed as follows: **5FA665, 5FA6651, 5FA6652, 5FA6653**, all listed with an **expiration date of April 17**.

Additional EPI Pen batch numbers were also confirmed:

Expiration date 4/2017: 5GN767, 5GN773, 5GM631;

Expiration date 5/2017: 5GM640

Expiration date 9/2017: 6GN215, 6GM082, 6GM072, 6GM081

Expiration date 10/2017: 6GM088, 6GM199, 6GM091, 6GM198, 6GM087

All epi-pen's stored in the health office have been checked and we are advising you of the recall so you can check your supply at home

If boxes feature any of the expiration dates or one of the aforementioned batch numbers, the company, Alphapharm, is encouraging consumers to replace the pens immediately by returning them to a pharmacist. The replacement of effected epi-pen will be at no cost to consumer.

If you have any questions, please contact your health care provider.

Thank you,

Gerald North
Superintendent of Schools